



**Al-Hami**  
Medical Equipment L.L.C



**BIOCHEMISTRY**



## Selectra Flexor Pro M

- 180 photometric tests/hour
- Sample Position: 50 Samples
- Reagent Position: 32 positions for
- Sample Volume : 1– 30  $\mu$ l in 0.1  $\mu$ l steps
- 8 different wavelengths between 340–700 nm

[www.alhamico.com](http://www.alhamico.com)