



**Al-Hami**  
Medical Equipment L.L.C



## HEMATOLOGY

### XT-1800i



- **Measurement principles:** Fluorescence Flow Cytometry
- **Aspiration volume :** **Close mode**150  $\mu$ L – **Manual mode**85  $\mu$ L
- **Throughput :** up to 80 samples/h
- **Parameters :** WBC, RBC, HGB, HCT, MCV, MCH, MCHC, PLT (Impedance and Fluorescent Optical) NEUT%, LYMPH%, MONO%, EO%, BASO%, NEUT#, LYMPH#, MONO#, EO#, ...

[www.alhamico.com](http://www.alhamico.com)