



Al-Hami

Medical Equipment LLC

Products Catalogue

Immunoassay



Al-Hami
Medical Equipment L.L.C



Abbott

IMMUNOASSAY



Alinity i

- **Method** : Chemiluminesce
- **Throughput**: 200 t/h
- **Sample Position**: 150 Samples
- **Reagent Position**: 47 positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C



Abbott

IMMUNOASSAY



Architect
i2000sr

- **Method** : Chemiluminescent
- **Throughput**: 200 t/h
- **Sample Position**: 135 Samples
- **Reagent Position**: 25positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C



Abbott

IMMUNOASSAY



Architect i1000sr

- **Method** : Chemiluminescent
- **Throughput**: 100 t/h
- **Sample Position**: 65 Samples
- **Reagent Position**: 25positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

SIEMENS
Healthineers

IMMUNOASSAY



Advia

Centaur XPT

- **Method** : Chemiluminescent
- **Throughput**: 240 t/h
- **Sample Position**: 180 Samples
- **Reagent Position**: 30 positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

SIEMENS
Healthineers

IMMUNOASSAY



Immulite 2000 XPI

- **Method** : Chemiluminescent
- **Throughput**: 200 t/h
- **Sample Position**: 90 Samples
- **Reagent Position**: 24positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

SIEMENS
Healthineers

IMMUNOASSAY



Advia Centaur CP

- **Method** : Chemiluminescent
- **Throughput**: 180 t/h
- **Sample Position**: 84 Samples
- **Reagent Position**: 24positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C



IMMUNOASSAY



Cobas e411

- **Method** : electrochemiluminescence (ECL)
- **Throughput**: 88 t/h
- **Sample Position**: 30 (Disk Version)- 75 (Rack Version)
- **Reagent Position**: 18 positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

Ortho
Clinical Diagnostics

IMMUNOASSAY



VITROS 3600

- **Method** : Chemiluminescent
- **Throughput**: 190 t/h
- **Sample Position**: 80 Samples
- **Reagent Position**: 31 positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

Ortho
Clinical Diagnostics

IMMUNOASSAY

VITROS ECiQ



- **Method** : Chemiluminescent
- **Throughput**: 90 t/h
- **Sample Position**: 60 Samples
- **Reagent Position**: 20positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C



**BECKMAN
COULTER**

IMMUNOASSAY

UniCel Dxl
600 Access



- **Method** : Chemiluminescent (CLIA)
- **Throughput**: 200 t/h
- **Sample Position**: 60 Samples
- **Reagent Position**: 50 positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C



**BECKMAN
COULTER**

IMMUNOASSAY

Access 2



- **Method** : Chemiluminescent
- **Throughput**: 100 t/h
- **Sample Position**: 60 Samples
- **Reagent Position**: 24 positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

mindray™

IMMUNOASSAY



CL-2000i

- **Method** : ALP-AMPPD method*
- **Throughput**: 240 t/h
- **Sample Position**: 300 Samples
- **Reagent Position**: 36positions

*helps realize high sensitivity measurements

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

mindray™

IMMUNOASSAY



CL-1200i

- **Method** : ALP-AMPPD method*
- **Throughput**: 180 t/h
- **Sample Position**: 60 Samples
- **Reagent Position**: 25positions

*helps realize high sensitivity measurements

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

mindray™

IMMUNOASSAY



CL-900i

- **Method** : ALP-AMPPD method*
- **Throughput**: 180 t/h
- **Sample Position**: 50 Samples
- **Reagent Position**: 15positions

*helps realize high sensitivity measurements

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

DiaSorin

IMMUNOASSAY



Liaison XL

- **Method** : chemiluminescent
- **Throughput**: Max. 180 t/h
- **Sample Position**: 120 sample
- **Reagent Position**: 25 positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C

DiaSorin

IMMUNOASSAY



Liaison

- **Method** : Flash chemiluminescent immunoassay
- **Throughput**: Max. 180 t/h
- **Sample Position**: 144 sample
- **Reagent Position**: 15 positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C



TOSOH BIOSCIENCE

IMMUNOASSAY



AIA 900

- **Method** : Biochromatic Fluorescence Kinetics
- **Throughput**: 90 t/h
- **Sample Position**: 100 Samples
- **Reagent Position**: up to 180 (9 Rack / 20 sample)

www.alhamico.com



Al-Hami
Medical Equipment L.L.C



TOSOH BIOSCIENCE

IMMUNOASSAY

AIA 360



- **Method** : Biochromatic Fluorescence Kinetics
- **Throughput**: 36 t/h
- **Sample Position**: 25 Samples
- **Reagent Position**: 25positions

www.alhamico.com



Al-Hami
Medical Equipment L.L.C



BIOMÉRIEUX

IMMUNOASSAY

VIDAS 3



- **Method** : Fluorescent Assay
- **Throughput**: 36 t/h
- **Sample Position**: 27 sample
- **Reagent Position**: 27 reagent

www.alhamico.com



Al-Hami
Medical Equipment L.L.C




BIOMÉRIEUX

IMMUNOASSAY

**VIDAS PC
& miniVidas**



- **Method** : Fluorescent Assay
- **Throughput**: 80 t/h – miniVidas 36 t/h
- **Sample Position**: 60 Samples - miniVidas 12 t/h
- **Reagent Position**: 60positions - miniVidas 12 t/h



Al-Hami

Medical Equipment LLC

www.alhamico.com



info@alhamico.com



+971-50 - 8507636